DISINFECTANT & FILTER CHANGE LOG DSD-201, SSD-102

SERIAL #: _____

DISINFECTANT CHANGE		DISINFECTANT FILTER (MF01-0011)		WATER FILTERS			WATERLINE SANITIZATION (Performed with every 0.2 micron filter change)		AIR FILTERS (MF01-0028)		CHARCOAL FILTER For Vapor Mgmt. (MF01-0037)
				1 MICRON (MF01-0014)	0.45 MICRON (MF01-0015)	0.2 MICRON (MF01-0013)					
Station A	Station B	Station A	Station B						Station A	Station B	
BASED ON MRC TEST OR HLD USE LIFE		CHANGE WITH EVERY DISINFECTANT CHANGE		PRESSURE BELOW 40 PSI OR EVERY 3 MONTHS		EVERY 6 MONTHS (MAX)	(CIRCLE ONE)		EVERY 3 MONTHS		EVERY 6 MONTHS
Date/Initials	Date/Initials	Date/Initials	Date/Initials	Date/Initials	Date/Initials	Date/Initials			Date/Initials	Date/Initials	Date/Initials
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			

www.minntech.com/medivators • Toll Free: 1-800-444-4729

Endoscope reprocessing the way it should be

MEDIVATORS®

© MINNTECH CORPORATION All rights reserved. 50096-116/ D