FILTER CHANGE LOG Advantage®

SERIAL # : _____

DISINFECTANT FILTER (MF01-0058)		BASIN DRAIN FILTER (MF01-0059)		WATER FILTERS			WATERLINE DISINFECTION (Performed with every 0.2 micron filter change)		AIR FILTER (MF01-0028)		CHARCOAL FILTER For Vapor Mgmt. (MF01-0064) On back of unit
				1 MICRON (MF01-0014)	.45 MICRON (MF01-0015)	0.2 MICRON (MF01-0013)					
LEFT	RIGHT	LEFT	RIGHT						LEFT (2)	RIGHT (2)	
EVERY DISINFECTANT CHANGE		EVERY DISINFECTANT CHANGE		PRESSURE RELATED BELOW 35 PSI OR EVERY 3 MONTHS		EVERY 6 MONTHS (MAX)	(CIRCLE ONE)		EVERY 6 MONTHS		EVERY 6 MONTHS
Date/Initials	Date/Initials	Date/Initials	Date/Initials	Date/Initials	Date/Initials	Date/Initials			Date/Initials	Date/Initials	Date/Initials
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			
							YES	NO			

© MINNTECH CORPORATION All rights reserved. 50092-520/A