

1. The scope retainer should be used on small diameter scopes to prevent the distal end from moving around during the disinfection cycle.
2. Position the scope in the basin and connect the hookup.
3. Slide the scope retainer over the distal end until approximately one to two inches of the tip are exposed. Refer to figure 1.



Figure 1